



Product Service

EC Certificate

EC Design-Examination Certificate

Directive 93/42/EEC on Medical Devices (MDD), Annex II (4)
(Devices in Class III)

No. G7 14 12 25701 058

Manufacturer: **B. Braun Surgical S.A.**
Ctra. de Terrassa, 121
08191 Rubi (Barcelona)
SPAIN

Product: **Sutures**

Model(s): **PremiCron**
PremiCron with PTFE Pledgets (6x3 mm, 3x3 mm)

Parameters: Sterile, synthetic, non-resorbable,
braided surgical suture with or without PTFE Pledgets
Diamter: USP 8/0 (0,4 metric) to
USP 5 (7 metric)
Length: max. 250 cm

The Certification Body of TÜV SÜD Product Service GmbH declares that a design examination has been carried out on the respective devices in accordance with MDD Annex II (4). The design of the devices conforms to the requirements of this Directive. For marketing of these devices an additional Annex II certificate is mandatory. See also notes overleaf.

Report no.: 713046480

Valid from: 2015-01-04

Valid until: 2020-01-03

Date, 2014-12-23

Hans-Heiner Junker



TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

Page 1 of 1