

Sterimedix Ltd

Quality System Document QD 2040 Issue P
EC Declaration of Conformity

Sterimedix Ltd. Declaration of Conformity for Vitreoretinal Cannulae

Ref: DOC11

30/05/2018 - Issue 10

Council Directive 93/42/EEC Concerning Medical Devices

The undersigned declares that the product named in this document meets the applicable Council Directive provisions that apply and the CE Mark may be affixed.

General Product Name Vitreoretinal Cannulae

Product Reference **Vitreoretinal Cannulae**
M4410; M4411; M4447; M4448; M4470; M4470C; M4471; M4477; M4478; M4478B;
M4478C; 03CA77; 03CA78; SD5205; SD4478
Heavy Liquid Infusion Aspiration Cannulae
M4450; M4451; M4452; M4453; M4454; M4473; M4474; M4499; M4499A; 03CA99;
03CA99A; 03CA52; 03CA53; SD4812; SD4806; SD4811; SD4474
Silicone Brush Cannulae
M4449; M4459; M4475; M4475A; SD4459; SD4805
Silicone Tip Cannulae
M4456; M4457; M4458; 03CA56; 03CA58; SD5229; SD5226; SD4807

Manufacturer Sterimedix Ltd
1 Madeley Road, North Moons Moat, Redditch
Worcestershire, B98 9NB, UK

Notified Body Notified Body 0120
SGS United Kingdom Ltd
202b Worle Parkway, Weston-super-Mare, BS22 6WA

Intended Use Surgically invasive devices for specific infusion and aspiration functions during
retinal procedures, and are intended for either transient or short-term use.

Sterile Sterile - EO. Approved by SGS United Kingdom Ltd

Measuring Function No

Conforming to Current Standards

EN 556-1	EN ISO 11607-1, 2
ISO 594-1, 2	EN ISO 11737-1, 2
BS EN 980	EN ISO 13485
EN 1041	EN ISO 14155-1
EN 1707	EN ISO 14937
EN ISO 10993-1, 3, 4, 5, 6, 7, 11	EN ISO 14971
EN ISO 11135-1	EN ISO 15223-1

Classification Class IIa in accordance with Annex IX Rule 6

CE Marking

Notified Body and Number:	SGS 0120
Initial Date of CE Marking:	21/05/1997
Certified to:	93/42/EEC
Certificate Number:	GB97/9964
Certificate Issue Date:	12/08/2015
Certificate Expiration Date:	06/02/2020

Assessment Route By Annex II excluding section 4 and Article 11 of the European Communities Council
Directive 93/42/EEC amended by Directive 2007/47/EC, and the application of BS
EN ISO 13485:2016.

This product does not contain human blood derivatives and is not manufactured utilising tissues of animal origin.

Signed  Date 31/5/18

Name: Andrew Ellis

Position: QA/RA Manager