



EC Declaration of Conformity

in Accordance with EC Directive 93/42/EEC on Medical Devices

Manufacturer Carl Zeiss Meditec AG, Goeschwitzer Strasse 51-52, 07745 Jena, Germany

We Carl Zeiss Meditec AG herewith declare with sole responsibility that the following Medical Device meets the Requirements of the European Directive 93/42/EEC. The device is provided with CE Marking.

Product identification:	<i>Intraocular lens</i>
Medical Device Trade Name:	<i>AT LISA tri toric 939MP</i>
Models/Reference:	<i>n/a</i>
Accessories:	<i>Including all options and accessories as defined in the User Manual</i>
Medical Device Class MDD 93/42/EEC:	<i>Class IIb</i>
Conformity Assessment Procedure	<i>Annex II of MDD 93/42/EEC</i>
Scope of Application:	<i>This Declaration of Conformity is valid for all products manufactured until 2021-11-29</i>
UMDNS classification:	<i>16-071</i>
GMDN Code	<i>35658</i>

Notified Body: DQS Medizinprodukte GmbH, August-Schanz-Straße 21, 60433 Frankfurt – notified under 0297.

Any Modification to the Product not authorized by Carl Zeiss Meditec AG will invalidate this Declaration.


i.V. Guillaume Gasc
Vice President Competence Center &
Order fulfillment consumables


i.V. Christina Jakob
Director Clinical & Global Regulatory Affairs
Implants / OVDs / Sterile Products

Berlin, 2016-11-22

Carl Zeiss Meditec AG, Berlin Site, Max-Dohrn-Str. 8-10; 10589 Berlin, Germany